RX50 Biological Microscope









- In order to correct the shortcoming of horizon guide rail, new stage is designed with double—way linear driving mechanism. This change protects the stage from overload at the end of both rails, improves the rigidity and performance of the stage.
- The handle of the stage can be set at each side based on users' preference. The X, Y biaxial adjustments are designed with low
- position for comfortable operation.

 Two slices can be hold on the stage by using damping-type double clips, easy for comparative study. Moving range: 80mm X55mm;
- Processed with special craft, the surface of the stage is anticorrosive and anti-friction. The platform with an arc transition design reduces the stress concentration and damage from impact.
- Light barrier, which can effectively reduce the stray light, improves the image contrast in fluorescence observation

Modular frame, improving the system compatibility

RX50 with modularization design, separated cross arm and main body, improves the system compatibility of biological and fluorescence frame.

Highly sensitive coaxial coarse and fine adjustment

Coaxial adjustment adopts double-stage driving, with adjustable tension tightness and upper limit stop, coarse range is 25mm and fine precision is $1\,\mu\,m$. Not only accurately focus but also precision measurement is available.

Viewing head with multi-splitting ratio

Viewing head with multi-splitting ratio
RX50 viewing head is designed of multiple options for splitting ratio.
with wide beam imaging system, 26.5mm super wide filed of view is available.
Trinocular head with inverted image, splitting ratio Binocular:
Trinocular=10.00 or 20.80 or 0:100. Except for concentrating 100% light to eyepiece tube or camera tube, there is another option with 20% light to eyepiece tube and 80% to camera tube, so that eyepiece observation and image output can be available at the same time.
Trinocular head with erect image, splitting ratio Binocular:Trinocular=100.0 or 0:100. The moving


- ☆ Quintuple/sextuple/septuple nosepiece for option.
 ☆ 12V100W halogen lamp house provides transmitted light.
 Illuminator center is preset, without adjustment after replacing the

lamp. $\times \text{NA.0.9 swing-out type achromatic condenser,with iris diaphragm and aperture scale, provides sufficient and uniform light for full-field observation in different magnifications.$









Multifunctional reflection illumination

In RX50 reflection fluorescent illumination, maximum 6 fluorescence filters can be assembled at the same time. Filters are placed in a rotary table for convenient switch. High precision and stable rotary table and high-performance imported filter ensures a

drift-free image.

There is a light shutter in front of the reflected illuminator.

- is linere is a light shutter in front of the fehicated illuminator. It is used to shut the fluorescent light to prevent fluorescence quenching of the slice.

 The light barrier can protect users from the harm of UV light.

 The use of ND attenuation filter, or aperture and field diaphragm rod can efficiently reduce the intensity of exciting light to protect. the slice.
- ☆ After replacing the lamp, the centering objective helps users in adjusting the filament center to make sure a sufficient and uniform

New digital power control system with operating time and current value, clearly shows the working state of the mercury lamp.

Two power supply systems, providing multiple choices of high quality illumination


New developed 100W EHV DC mercury lamp house with improved thermal cycle greatly reduces the surface temperature of the lamp house and avoids the scald risk during operation. The filament

75W xenon lamp for option.





Standard parameters of fluorescent filters, substantially meet the needs of conventional scientific research. For some special requirements, specific filters also can be provided corresponding to fluorescent probes.



RX50 Specifications

Optical system	Infinity color corrected optical system				
	Inverted image, 30° inclined gemel trinocular head, interpupillary distance: 50mm~76mm; splitting ratio R:T=100:0 or 20:80 or 0:100				
	Erect image, 30° inclined gemel trinocular head, interpupillary distance: 50mm~76mm; splitting ratio R: T=100:0 or 0:100				
	High eyepoint wide field plan eyepiece PL10X25mm, diopter adjustable				
	High eyepoint wide field plan eyepiece PL10X25mm, with reticle, diopter adjustable				
	High eyepoint wide field plan eyepiece PL10X26.5mm, diopter adjustable				
	High eyepoint wide field plan eyepiece PL10X26.5mm, with reticle, diopter adjustable				
Objective	Plan semi-apochromatic fluorescence objectives (4X/10X/20X/40X/100X)				
Nosepiece (with DIC slot)	Quintuple nosepiece				
	Sextuple nosepiece				
	Septuple nosepiece				
	Biological frame (transmitted), low-position coaxial coarse and fine adjustment, coarse adjustment distance: 25mm; fine precision: 0.001mm, With coarse adjustment stop and tightness adjustment Built-in 100-240V_AC50/60Hz wide voltage transformer; intensity adjustable by digital set and reset; built-in transmitted filters LBO/ND6/ND25.				
	Fluorescence frame (transmitted), low-position coaxial coarse and fine adjustment, coarse adjustment distance: 25mm; fine precision: 0.001mm; With coarse adjustment stop and tightness adjustment.Bull:-in 100-240V_AC50/80Hz wide voltage transformer, intensity adjustable by diatial set and reset; built-in transmitted filters LBo/ND6/ND25).				
Stage	Double layers mechanical stage, size: 187mm X168mm; moving range: 80mm X55mm; precision: 0.1mm; two-way linear drive tension adjustable				
Condenser	Swing-out type achromatic condenser (N.A.0.9)				
Reflected fluorescence illuminator	Sextuple reflected fluorescence illuminator with iris field diaphragm and aperture diaphragm, central adjustable; with filter slot and polarizing slot; with fluorescence filters (UV/B/G for option).				
	100W mercury lamp house, filament center and focus adjustable; with reflected mirror, mirror center and focus adjustable. (75W xenon lamp house for option)				
	Digital power controller, wide voltage 100–240VAC				
	Imported OSRAM 100W mercury lamp.(OSRAM 75W xenon lamp for option)				
lamp house	12V/100W halogen lamp house for transmitted light, center pre-set, intensity adjustable				
Other accessories	Camera adapter: 0.5X/0.65X/1X focusing C-mount				
	Cooled CCD camera, SONY2/31.4M ICX285AQ COLOR CCD				
	Centering objective for fluorescence observation				
	Professional software for fluorescence analysis				
	High precision micrometer, scale value 0.01mm				





Professional plan fluorite objectives
PLAN-FLUOR series infinity plan semi-apochromatic objectives, is the best choice for
PLAN-FLUOR series infinity plan semi-apochromatic objectives, is the best choice for
plantage or professional professional operation patients, perfectly corrects all kinds
of chromatic aberrations. Lurge numerical aperture design presents high resolution, high
contrast micro-images, in fluorescence observation, the image is clear and bright while
the background is pure black, prominent for ultraviolet fluorescence.

		Numerical Aperture	Working Distance	Cover-glass Thickness		Spring
	4X	0.13	16.43	0.17	- /	- /
Plan semi-	10X	0.30	8.13	0.17	1	1
apochromatic	20X	0.50	2.03	0.17	1	1
objective	40 X	0.75	0.74	0.17	1	Spring
	100X	1.28	0.18	0.17	Oil	Spring

- High eye-point ultra wide field plan eyepieces

 Breaking through 22mm conventional field of view, reaching to 25mm and 26 5mm, much smoother and vider field of view, is helpful to improve the vort efficiency.


 Locaring pin on eyepiece insorts fro the eyepiece table, fiving the eyepiece for easy focusing Laguer adjustation grand of depter from 1-8 bit, meter more demand of different ruses.

 Eyepiece cup can be turn up to avoid external stray light. Spectacle-vearers should turn down the cup to protect both spectacles and eyepieces.

			Diopter Adjustment	
High eye-point wide field plan eyepiece	10X	25	-5 ~ +5	-
			Diopter Adjustment	
High eye-point wide	10X	26.5	-8 +5	

09/10

RX50 System diagram: mm



RX50 Dimension: mm

